



TopWeigh Ltd

Scale House, (Jeffrey Estates), Rockcliffe,
Carlisle, Cumbria. CA6 4BH. UK



Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
Mobile :+44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate

Customer Cumbria Kart Racing Club		Customer Order No.		26 Jan 2019	
Site Address :- Rowrah Cumbria		Date of Last Test		8 Feb 2020	
		Date of Test		8 Feb 2021	
		Next Test Due		8 Feb 2021	
		Next Test Due within 30 days of above.			
		Machine		Platform	
		Type			
		Manufacturer		TopWeigh Ltd	
		Capacity :- Max		300 kg	
		Min			
		e.		0.1 k	
Calibration Certificate No.		TW	1020100	Maximum Tolerance	
Initial Test		Calibration Test		Serial No.	
				JWS-06-300	
Weights Applied	Reading	Weights Applied	Reading	Class	
				(1	
Observations & Comments :-					
Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 Found to be CORRECT					
Weights Traceable to National Standard					
TW 10796					
Weights Used TWR Set					

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :- *A Jeffrey*

Print Name :- A J Jeffrey