



**SUNBELT RENTALS**  
Unit 5, Malmo Court  
Kirkleatham Business Park  
Redcar  
Cleveland  
TS10 5SQ

D.Harbinson  G.Greenfield   
A.Harbinson  L.Spence   
T.Tate  T.Gair

Page 1 of 2  
Certificate Number :  
**SBR3554**  
Approved Signatory

pp

Tel: 0370 330 6021

Email: [Instrumentation@sunbeltrentals.co.uk](mailto:Instrumentation@sunbeltrentals.co.uk)

# CERTIFICATE OF CALIBRATION

**Customer :** LARKHALL CIRCUIT  
MERYTON ROAD  
  
LARKHALL  
ML9 2UL

**Instrument :** System ID : ID17403  
Description : ACOUSTIC CALIBRATOR  
Manufacturer : CASELLA  
Model Number : CEL-282  
Serial Number : 2/11820058  
Procedure Number : 3163  
  
Customer Ref :  
Location :

**Environmental Conditions :**

Temperature : 20°C ± 0°C  
Relative Humidity : 50%RH ± 50%RH  
Mains Voltage : 240V ± 30V  
Mains Frequency : 50Hz ± 0Hz

**Comments :**

**Traceability Information :**

<i>Instrument Description</i>	<i>Serial Number</i>	<i>Certificate Number</i>	<i>Cal. Date</i>	<i>Cal. Period (Weeks)</i>
CIRRUS C	B17862FD; 20042787	H14077	19/04/21	52

Calibrated By : **Trevor Tate**

Date of Calibration : **24/06/2021**  
Calibration Due Date : **23/06/2022**

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised National Standards, and to the units of measurements realised at the National Physical Laboratory or other recognised National Standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



# CERTIFICATE OF CALIBRATION

Certificate Number :  
SBR3554

Test Title	Applied Value	Tolerance	Reading	Pass / Fail
<b>SOUND PRESSURE LEVEL OUTPUT</b>				
114dB Output	113.7dB	+/- 500mdB	113.52dB	PASS
<b>End of Test</b>				

*END OF TEST DATA*